

To:

**THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES**  
**Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)**  
**Director General of Engineering Affairs, MOH**  
**Director General of Royal Hospital**  
**Director General of Khoula Hospital**  
**Director General of Medical Supplies (MOH)**  
**Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)**  
**Hospital Director (Al Nahda Hospital)**  
**Hospital Director (Al Massara Hospital)**  
**The Head of Medical Services in SQU Hospital**  
**The Head of Medical Services in Royal Oman Police**  
**The Head of Medical Services in Ministry of Defence**  
**The Head of Medical Services in The Diwan**  
**The Head of Medical Services in The Sultan's Special Force**  
**The Head of Medical Services in Internal Security Services**  
**The Head of Medical Services in Petroleum Development of Oman**  
**The Head of Medical Services in LNG Oman**  
**ALL PRIVATE PHARMACIES & DRUG STORES**


After Compliments,

Please find attached our Circular No. 184... dated 27/10/2021 Regarding GHC Recall of DLP® Left Heart Vent Catheters from ( mrf: Medtronic).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 184/2021


بنقدف بثقة  
Moving Forward  
with Confidence

رؤية عُمان  
2040  
Oman Vision

20 -03-1443 H

27 -10-2021

**Recall of DLP® Left Heart Vent Catheters from Medtronic.**

Source	GHC- Gulf Health Council.
Product	DLP® Left Heart Vent Catheters.
Manufacturer	Medtronic.
Local Agent	Alzahrawi Medical Supplies LLC.
The affected products	Product Name: DLP® Left Heart Vent Catheters, Model Number: 16 Fr 12116, 18 Fr 12118. Product Name: DLP® Left Heart Vent Catheters, Model Number: 18 Fr 12118.
Reason	Medtronic is voluntarily recalling specific models of unused above-mentioned products, due to the potential for a wire protrusion through the left heart vent catheter tip. If unnoticed prior to the procedure, this wire protrusion could lead to tissue damage (abrasion/perforation) and thereby to a longer duration of the procedure and/or surgical repair. In those procedures where a Model 12116 or Model 12118 was used, monitor patients in the acute post-op care setting for bleeding.
Action	1. Do not use any affected products, and return all unused affected product. 2. Contact the local agent for remedial action.
Product Image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control contact E-mail: <a href="mailto:Med-device@moh.gov.om">Med-device@moh.gov.om</a>



Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL