

# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control  
MUSCAT



سِلاطِنَا عُمَانِ  
وَزَارَةُ الصِّحَّةِ  
وَالرِّقَابَةِ الدَّوَلِيَّةِ  
مَسْقَط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES  
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)  
Director General of Engineering Affairs, MOH  
Director General of Royal Hospital  
Director General of Khoula Hospital  
Director General of Medical Supplies (MOH)  
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)  
Hospital Director (Al Nahda Hospital)  
Hospital Director (Al Massara Hospital)  
The Head of Medical Services in SQU Hospital  
The Head of Medical Services in Royal Oman Police  
The Head of Medical Services in Ministry of Defence  
The Head of Medical Services in The Diwan  
The Head of Medical Services in The Sultan's Special Force  
The Head of Medical Services in Internal Security Services  
The Head of Medical Services in Petroleum Development of Oman  
The Head of Medical Services in LNG Oman  
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...<sup>50</sup>..... dated <sup>02/10/20</sup> Regarding recall of Skin Preparation  
Electrode gel LemonPrep, PediaPrep, Wave Prep and Cardio Prep from Mavidon Medical

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control

MUSCAT

Circular No. 50 / 2020

07 -07-1441 H

02 -03-2020

Ref: 37/2020



سِلاطِنَاةُ عُومَانِ  
وَزَارَةُ الصِّحَّةِ  
وَالرِّقَابَةِ الدَّوَلِيَّةِ  
مَسْقَط

## Recall of Skin Preparation Electrode gel LemonPrep, PediaPrep, Wave Prep and Cardio Prep from Mavidon Medical

Source of Recall	Health Gulf Council
Product	Skin Preparation Electrode gel LemonPrep, PediaPrep, Wave Prep and Cardio Prep
Manufacturer	Mavidon Medical
The affected products	All lots of LemonPrep, PediaPrep, Wave Prep and Cardio Prep
Reason for Recall	products may be contaminated with the microorganism Burkholderia cepacia leading to an infection risk to patient
Action	1. Kindly check your stock, contact your local agent for remedial action
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail <a href="mailto:dg-padc@moh.gov.om">dg-padc@moh.gov.om</a>

Directorate General of Pharmaceutical Affairs & Drug Control  
Sultanate of Oman

Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL

