

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للصحة
والسرعة والدواء
مسقط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. 231..... dated 09/12/2028 Regarding NCMDR Recall of Spinal Rod Cutter from (mfr: Zimmer Biomet).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control
MUSCAT




سلطنة عمان
وزارة الصحة
المديرية العامة للأجهزة
والأدوية الطبية
مسقط

Circular No. 231/2020

23 -04-1442 H

09 -12-2020

Recall of Spinal Rod Cutter from Zimmer Biomet.

Source	NCMDR - National Centre Medical Device Reporting- SFDA. https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=15455
Product	Spinal Rod Cutter Maximum Pin Diameter ¼ Inch (6.4M).
Description	Spinal Rod Cutter.
Manufacturer	Zimmer Biomet.
Local agent	Surgitech Equipments L.L.C.
The affected products	Item Number: 00-3925-002-00 Lot Number: All Lots.
Reason	A potential for fracture during use.
Action	1. Immediately locate and quarantine all affected products in your inventory, after that return them to Zimmer Biomet. 2. Contact the local agent for remedial action.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through an E-mail: Med-device@moh.gov.om



Dr. Mohammed Hamdan Al Rubaie

Director General