


Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سِلاطِنَة اعمان
وزارة الصحة
للمديرية العامة للصحة
والرقابة الدوائية
مسقط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...86..... dated 22.10.12 Regarding NHRA recall of Examination Gloves (Mfr:Lotus)

Copy to:


- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للأدوية
والرقابة الدوائية
مسقط

Circular No. 86 / 2020

28 -Shabaan-1441 H

22 -April-2020

Recall of Examination Gloves from Lotus – China

Source of Recall	الهيئة الوطنية لتنظيم المهن والخدمات الصحية (NHRA)
Product	Examination Gloves
Manufacturer	Lotus – China
The affected products	All batches
Reason for Recall	NHRA (Bahrain) initiated this recall due the material used in the above mentioned gloves is being used before which may affect the sterility of the gloves
Action	<ul style="list-style-type: none">Stop using the above product.Return the available stock if any to the local agent.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail dg-padc@moh.gov.om

Directorate General of Pharmaceutical Affairs & Drug Control
Sultanate of Oman



Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL