


Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control
MUSCAT



سِلْطَنَةُ عُومَانِ
وَزَارَةُ الصِّحَّةِ
وَالرِّقَابَةِ الدَّوَالِيَّةِ
مَسْقَط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. ^{166/2020} dated ^{3/9/2020} Regarding FDA Recall of Cytocell LPH533-A NUP98 Distal Probe Green from (mfr: Cytocell LTD).


Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control
MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للصحة
والرقابة الدوائية
مسقط

Circular No. 166 / 2020

14 -01-1442 H

03 -09-2020

Recall of Cytocell LPH533-A NUP98 Distal Probe Green from Cytocell LTD.

Source	FDA-Food and Drug Administration. https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRES/res.cfm?id=182622
Product	Cytocell LPH533-A NUP98 Distal Probe Green.
Description	Reagents, specific, analyte.
Manufacturer	Cytocell LTD.
The affected products	Lot numbers: 069915, 070433, 069305, 069468
Reason	A low risk of a false positive result being issued with a laboratory developed test (LDT) that utilizes the LPH533-A NUP98 Distal Probe Green.
Action	<ol style="list-style-type: none">1. Immediately examine your inventory and quarantine all product subject to recall.2. Cytocell requests that you destroy the remaining inventory. Please complete and return the enclosed response form as soon as possible.3. Contact the local agent for remedial action.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through an E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL

