

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سِيَاظَتَا مَمْلَاكَا
وَزَارَاةَا الصِّحَاةَا
اَلْمَدِيرِيَّةَا الْعَامَّةَا لَلصِّدَاكَاةَا
وَالرَّقَابَاةَا الدَّوَالِيَّةَا
مَسَقَطَا

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...185... dated 08/10/20 Regarding NCMDR Recall of Isolation Gowns from (mfr: Xiantao Fushi Protective products CO., LTD).

Copy to:


- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT




سلطنة عمان
وزارة الصحة
المديرية العامة للأجهزة الطبية
والرقابة الدوائية
مسقط

Circular No. 185/2020

20 -02-1442 H

08 -10-2020

Recall of Isolation Gowns from Xiantao Fushi Protective products CO., LTD.

Source	NCMDR - National Centre Medical Device Reporting- SFDA. https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=4&rid=15373
Product	Isolation Gowns.
Description	Gowns, Medical, Disposable.
Manufacturer	Xiantao Fushi Protective products CO., LTD.
The affected products	Lot Number: FS19264 Expiration Date: 11-2024
Reason	The affected gowns are easy to tear when wearing.
Action	Contact the supplier to replace the items.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through an E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie

Director General

