

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سِلاطِنَا اَمَانَا
وَزَارَةُ الصِّحَّةِ
وَالشَّرِيعَةِ الْعَامَّةِ لِلصِّدْقِ
وَالسَّرَاقِبَةِ الدَّوْلِيَّةِ
مَسْقَط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. 54..... dated 03/03/12 regarding recall of Alinity ci-series System Control Module (SCM) from Abbott

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs
and Drug Control

MUSCAT



سلطنة عمان
وزارة الصحة
المديرية العامة للصحة
والرقابة الدوائية
مسقط


Circular No. 54 / 2020

08 -07-1441 H

Ref: 41/2020

03 -03-2020

Recall of Alinity ci-series System Control Module (SCM) from Abbott

Source of Recall	NCMDR National Centre for Medical Devices Reporting, SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=15054
Product	Alinity ci-series System Control Module (SCM)- General laboratoryware IVD
Manufacturer	Abbott
Local Agent	Waleed Pharmacy
The affected products	List Number: 03R70-01 Serial Number: All UDI: N/A
Reason for Recall	Abbott has identified potential performance issues for the Alinity ci-series Software version 2.6.2 and earlier.
Action	1. Kindly check your stock, contact your local agent for remedial action 2. Abbott is releasing Alinity ci-series Software version 3.1.0 to correct these issues.
Product image	
Comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail dg-padc@moh.gov.om

Directorate General of Pharmaceutical Affairs & Drug Control
Sultanate of Oman


Dr. Mohammed Hamdan Al Rubaie
DIRECTOR GENERAL

