



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 221 dated 29/11/2022 Regarding NCMDR Field Safety Corrective Action of Syngo.via with VB20, VB30, VB40, VB60 and VB70 software from (mfr: SIEMENS).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information



Circular No. 221/2022

5 -05-1444 H

29-11-2022

نتقدم بثقة
Moving Forward
With Confidence

رؤية عُمان
2040
Oman Vision

Field Safety Corrective Action of Syngo.via with VB20, VB30, VB40, VB60 and VB70 software from SIEMENS.

Source	NCMDR- National Center for Medical Devices Reporting- SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=10&rid=17352
Product	Syngo.via with VB20, VB30, VB40, VB60 and VB70 software.
Description	Radiology DICOM image processing application software.
Manufacturer	SIEMENS.
Local agent	Muscat Pharmacy & Stores LLC.
The affected products	VB20, VB30, VB40, VB60 and VB70 software.
Reason	In case: 1. Enhanced MRI images with rectangular FoV (Field of View) are created at the MR scanner and sent to syngo.via, and 2. at the syngo.via system, the user rotates the images to display the images in arbitrary orientation and saves them, and 3. afterwards, the user exports the images to another DICOM node using MR Export Rules (interoperability mode) then distance measurements on the newly created series might be wrong when being performed at systems that receive these new series.
Action	1. Customers are advised to follow the workaround steps outlined in the customer letter (supplied to impacted facilities). 2. Siemens will solve the issue with a software update which is now available. 3. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie
Director General

