


# Sultanate of Oman

Ministry of Health

Directorate General of Pharmaceutical Affairs  
and Drug Control  
MUSCAT



سِيَّاطِنَا مُمَا  
وَزَارَةُ الصِّحَّةِ  
الْمَدِيرِيَّةُ الْعَامَّةُ لِلصِّدْقِ  
وَالرَّقَابَةِ الدَّوَلِيَّةِ  
مَسْقَط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...٥٩..... dated 28/10/20 Regarding Anios laboratoires through Trade & Marketing Professionals Recall of Surf Safe premium (Mfr: Anios laboratoires)

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information

# Sultanate of Oman

Ministry of Health


Directorate General of Pharmaceutical Affairs  
and Drug Control

MUSCAT

Circular No. 09 / 2020

24-051441 H

20 -01-2020



سلطنة عمان  
وزارة الصحة  
المديرية العامة للصيدل  
والرقابة الدوائية  
مسقط

Ref: 4/2020

## Recall of Surf Safe premium from Anios laboratoires

Source of Recall	Anios laboratoires through Trade & Marketing Professionals
Product	1. BACCIDE KIT PRO 2. ECI MND PREMIUM 3. INTRUNET SURFA SAFE PREMIUM 4. SURFA'SAFE PREMIUM
Manufacturer	Anios laboratoires
Local Agent	Trade & Marketing Professionals
The affected products	<a href="http://www.anios.com/communication/EN/FSN_ANIOS_SSP_DISTRIBUTORS_EN_EX%20EU.pdf">http://www.anios.com/communication/EN/FSN_ANIOS_SSP_DISTRIBUTORS_EN_EX%20EU.pdf</a>
Reason for Recall	Anios laboratoires is recalling all batches as they do not comply with the quality expectations. They may contain the opportunistic environmental microorganism Burkholderia Cepacia
Action	1. Kindly check your stock, contact your local agent for remedial action 2. quarantine the affected Batches
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical Device to Director of Medical Device Control contact E-mail <a href="mailto:dg-padc@moh.gov.om">dg-padc@moh.gov.om</a>

Directorate General of Pharmaceutical Affairs & Drug Control  
Sultanate of Oman

Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL

