

بمقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040
Oman Vision

To:


THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Spécial Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No. 37..... dated 17/2/22 Regarding NCMDR recall of VITROS Immunodiagnostic Products HIV Combo Reagent Pack from (mrf: Ortho-Clinical Diagnostics).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information



Circular No. 37/2022


16 -07-1443 H


17 -02-2022

تقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040

Recall of VITROS Immunodiagnostic Products HIV Combo Reagent Pack from Ortho-Clinical Diagnostics

Source	NCMDR-National Center for Medical Device Reporting https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=16016
Product	VITROS Immunodiagnostic Products HIV Combo Reagent Pack.
Description	IVD.
Manufacturer	Ortho-Clinical Diagnostics.
Affected	Product Code (UDI): 6842779 (10758750031061) (xUS/CA) Affected Lot Numbers: 0520, 0530, 0540, 0570, 0580; Expiry: Expired 0590; Expiry: 07MAR2022 0630; Expiry: 11JUN2022 0640; Expiry: 18JUL2022
Local Agent	Al Hashar Pharmacy.
Reason	False Repeat Reactive Results.
Action	1. If customers have an affected lot that is still within expiration date, they may request replacement/credit. 2. Contact the local agent for remedial action.
Product Image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control contact E-mail: Med-device@moh.gov.om


Dr. Mohammed Hamdan Al Rubaie
DIRECTOR GENERAL

