



نتقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040
Oman2040

To:
THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 161 dated 25/11/24 Regarding SFDA Recall of Aggressive Cutter, Full Radius Resector, Aggr. Pro Line Shaver Blade, Round Burr, Aggressive Barrel Burr, Finish Barrel Burr, Semi Hooded Barrel Burr, Aggressive Barrel Burr, sterile from (mfr: Karl Storz SE & Co. Kg).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DSC
- Director of Pharmacovigilance & Drug Information Dept, DSC
- Director of Drug Control Department, DSC
- Director of Pharmaceutical Licensing Department, DSC
- Director of Central Quality Control Lab., DSC
- Supdt. of Central Drug Information



Circular No. 161 / 2024

نتقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040
Oman Vision

23 -05-1446 H
25 -11-2024

Recall of Aggressive Cutter, Full Radius Resector, Aggr. Pro Line Shaver Blade, Round Burr, Aggressive Barrel Burr, Finish Barrel Burr, Semi Hooded Barrel Burr, Aggressive Barrel Burr, sterile from Karl Storz SE & Co. Kg.

Source	SFDA- Saudi Food & Drug Authority. https://ade.sfda.gov.sa/Fsca/PublishDetails/167
Product	Aggressive Cutter, Full Radius Resector, Aggr. Pro Line Shaver Blade, Round Burr, Aggressive Barrel Burr, Finish Barrel Burr, Semi Hooded Barrel Burr, Aggressive Barrel Burr, sterile.
Manufacturer	Karl Storz SE & Co.Kg.
Local agent	Mustafa Sultan Science & Industry Co.LLC.
The affected products	Artikel (Name): 28208BKS (Aggressive Cutter, sterile) 28206CBS (Full Radius Resector, sterile) 28205NDS (Aggr. Pro Line Shaver Blade, sterile) 28205FDS (Round Burr, sterile) 28205HES (Aggressive Barrel Burr, sterile) 28205GDS (Finish Barrel Burr, sterile) 28208IDS (Semi Hooded Barrel Burr, sterile) 28205HDS (Aggressive Barrel Burr, sterile) Please refer to the attachment for the affected LOT numbers.
Reason	It was found that there are holes in the sterile barrier system.
Action	1. Immediately quarantine and discontinue use of associated part numbers listed in the attachment. 2. Contact the local agent to return the affected products.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: vigilance-md@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie
Director General



Rev 1: October 2024

FSN Ref: 24-0004

FSCA Ref: PFA-24-0004

Date: 14/10/2024

Urgent Field Safety Notice Product RECALL

28208BKS	Aggressive Cutter, sterile
28206CBS	Full Radius Resector, sterile
28205NDS	Aggr. Pro Line Shaver Blade, sterile
28205FDS	Round Burr, sterile
28205HES	Aggressive Barrel Burr, sterile
28205GDS	Finish Barrel Burr, sterile
28208IDS	Semi Hooded Barrel Burr, sterile
28205HDS	Aggressive Barrel Burr, sterile

For Attention of: Representatives for medical product safety, users, operators, distributors

Unique Device Identifier (s) (UDI-DI) :	n/a
Affected serial or lot numbers:	See table in attachment
FSN Type:	1 st Rev.

I. Identification of Affected Devices

The medical devices are suitable for use in minimally invasive investigations and treatments of a joint such as knee joints, shoulder joints, hip joints, small and medium joints (such as elbows, wrists, and ankles). Shaverblades are intended to remove tissue/bone. Shaverblades are surgically invasive and meant for transient use.

II. Reason for the Field Safety Corrective Action (FSCA)

a. Description of the product problem

It was found that there are holes in the sterile barrier system. This issue affects the attached lot numbers of the referenced KARL STORZ article numbers.

b. Background of the issue

During the update of the technical documentation, it was determined that there are holes in the sterile barrier system; due to the compromised sterile packaging, the affected products are being recalled.

c. Hazard giving rise to the FSCA

Due to the compromised sterile packaging, there is an increased risk of the patient being exposed to an infection. The use of the above-mentioned products should be discontinued.

d. Risks to patient/user or third parties

The use of one of the affected products carries the risk of infection for the patient. There is no further risk for the patient or user.

e. Other information relevant to FSCA

To date, no incidents have been reported to KARL STORZ in connection with the above-described issue – the corrective action (RECALL) is a preventive measure.

III. Type of Action to mitigate the risk

a. Action to be taken by the user

1. Immediately quarantine and discontinue use of associated part numbers listed.
2. Pass on this urgent field safety notice to all users of the products listed above and all other persons who need to be aware within your organization.
3. If you have or may have distributed the devices listed, please identify and promptly notify those recipients, or provide KARL STORZ a list of customers who received/may have received the products listed.
4. Return the filled feedback form by Fax or E-Mail to the indicated contact within 15 calendar days from the date of receipt.
5. Get in touch with your KARL STORZ representative to return affected products.
6. Please report any incidents related to this issue to the manufacturer, dealer or local representative and, if applicable, to the national competent authority, as this is important feedback.

Related to this action, no specific follow-ups on patients who have already been treated with products affected are required.

b. Action Being Taken by the Manufacturer

Recall of the affected products.

Artikel	Name	Alternative
28205HES	Aggressive Barrel Burr, sterile	28205HE - Aggressive Barrel Burr
28205HDS	Aggressive Barrel Burr, sterile	28205HD - Aggressive Barrel Burr

Please return the completed (signed and stamped) reply form within 15 calendar days from the Date of receipt.

Contact details of local representative (name, e-mail, telephone, address). This could be a distributor or KS subsidiary.

Name:

Telephone:

E-Mail:

The Competent (Regulatory) Authority of your country has been informed about this communication to customers.

On behalf of KARL STORZ, we thank you for your help and apologize for any inconvenience.

Yours sincerely,

KARL STORZ SE & Co. KG

i. V. Karim Djamshidi
Vice President
Global Patient Health & Regulatory Compliance

This document was created electronically and is valid without signature

