



Circular No. 18 / 2025

نتقدم بـ  
Moving Forward  
with Confidence

رؤية عمان  
2040

29 -07-1446 H  
29 -01-2025

Field Safety Notice of ROTEM sigma complete & ROTEM sigma complete + hep from Tem Innovations GmbH (Werfen).

Source	SFDA- Saudi Food & Drug Authority <a href="https://ade.sfda.gov.sa/Fsca/PublishDetails/250">https://ade.sfda.gov.sa/Fsca/PublishDetails/250</a>
Product	ROTEM sigma complete & ROTEM sigma complete + hep.
Manufacturer	Tem Innovations GmbH (Werfen).
Local agent	Muscat Pharmacy & Stores LLC.
The affected products	ROTEM sigma complete. - Part No.: 555501 - Lot No.: S221203 ROTEM sigma complete + hep. - Part No.: 555502 - Lot No.: S221204 Exp. Date (YYYY-MM-DD): 2024-06-30.
Reason	The FIBTEM C A5 results obtained with the current IVDR version exhibited a bias from measurements made with the previous IVDD version for a population within the Obstetric setting, leading to an adjustment in a bleeding management algorithm.
Action	The manufacturer instructs the customers to follow these actions: 1. Check if you have established a bleeding management algorithm for ROTEM sigma, 2. If yes, verify the established bleeding management algorithm for a potential impact due to the bias between the IVDD and the IVDR cartridges and take appropriate validation actions for your algorithm following your internal procedures. 3. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: <a href="mailto:vigilance-md@moh.gov.om">vigilance-md@moh.gov.om</a>

Dr. Mohammed Hamdan Al Rubaie  
Director General





نتقدم بثقة  
Moving Forward  
with Confidence

رؤية عمان  
2040  
Oman Vision

To:

**THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES**  
**Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)**  
**Director General of Engineering Affairs, MOH**  
**Director General of Royal Hospital**  
**Director General of Khoula Hospital**  
**Director General of Medical Supplies (MOH)**  
**Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)**  
**Hospital Director (Al Nahda Hospital)**  
**Hospital Director (Al Massara Hospital)**  
**The Head of Medical Services in SQU Hospital**  
**The Head of Medical Services in Royal Oman Police**  
**The Head of Medical Services in Ministry of Defence**  
**The Head of Medical Services in The Diwan**  
**The Head of Medical Services in The Sultan's Special Force**  
**The Head of Medical Services in Internal Security Services**  
**The Head of Medical Services in Petroleum Development of Oman**  
**The Head of Medical Services in LNG Oman**  
**ALL PRIVATE PHARMACIES & DRUG STORES**

After Compliments,

Please find attached our Circular No 18/2025 dated 29/11/2025 Regarding SFDA Field Safety Notice of ROTEM sigma complete & ROTEM sigma complete + hep from (mfr: Tem Innovations GmbH (Werfen)).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DSC
- Director of Pharmacovigilance & Drug Information Dept, DSC
- Director of Drug Control Department, DSC
- Director of Pharmaceutical Licensing Department, DSC
- Director of Central Quality Control Lab., DSC
- Supdt. of Central Drug Information



**DSC**  
مركز سلامة الدواء  
Drug Safety Center



**URGENT FIELD SAFETY NOTICE**  
**ROTEM® sigma complete Part No. 555501**  
**ROTEM® sigma complete + hep Part No. 555502**

November 15, 2024

Dear Valued Customer:

This notification is intended to advise your facility regarding a performance issue identified with ROTEM sigma complete and ROTEM sigma complete + hep product lot numbers starting with the letter "S" + 5-digit number from 221203 (ROTEM sigma complete) resp. 221204 (ROTEM sigma complete + hep). The table below identifies the initial LOT number per product.

Product Name	Part No.	Lot No.	Exp. Date(YYYY-MM-DD)
ROTEM sigma complete	555501	S221203	2024-06-30
ROTEM sigma complete + hep	555502	S221204	2024-06-30

**• Issue Description and Impact**

It has come to our attention through a customer complaint, that the FIBTEM C A5 results obtained with the current IVDR version exhibited a bias from measurements made with the previous IVDD version for a population within the Obstetric setting, leading to an adjustment in a bleeding management algorithm.

**• Mandatory Customer Actions**

Based on the above, please take the following **immediate** actions:

- **Check** if you have established a bleeding management algorithm for ROTEM sigma
- **If yes, verify** the established bleeding management algorithm for a potential impact due to the bias between the IVDD and the IVDR cartridges and take appropriate validation actions for your algorithm following your internal procedures.
- **Document** the acknowledgement on the Customer Reply Form and **return** the completed and signed form to the fax number or e-mail address listed on the next page.
- **Share** this information with your laboratory staff and follow your internal procedures.
- **Forward** this notification to all affected locations within your facility.
- **Retain** a copy of this notification for your records.

**• Customer Reply Form**

The customer reply form can be communicated to Werfen via the below options:

- e-mail address: tem-ra@werfen.com
- Fax no.: + 49-89-45429522

**• Contact information for questions**

- For technical questions please contact your local Werfen representative

We appreciate your prompt attention to this Urgent Field Safety Notice Letter.

Sincerely,

Signed by David Jacob  
 | I approve this document  
| 11/15/2024 | 3:52:00 PM CET  
599C1978CA4E43599FB2C4DDA08F4347

David Jacob

Director of Quality Assurance and Regulatory Affairs, PBM  
PRRC  
Tem Innovations GmbH