



Circular No. 23 / 2023


07-07-1444 H

29-01-2023

نتقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040
Oman Vision

Field Safety Corrective Action of DR-XD 1000(FDR Nano) from FUJIFILM Medical Systems

Source	NCMDR- National Center for Medical Devices Reporting- SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=6&rid=18417
Product	DR-XD 1000(FDR Nano).
Description	Radiological technology - mobile radiological diagnostic facilities.
Manufacturer	Fujifilm Medical Systems.
Local Agent	Advanced International Business Group (AIBG).
The affected products	DR-XD1000 Serial No. with last 4 numbers 1993 or less, see attachment for detailed instructions to identify the Serial Number of the DR-XD1000.
Reason	The arm part wobbled while moving this device and that an arm broke during arm movement.
Action	1. FUJIFILM will correct and implement the measures. 2. Contact the local agent for remedial action.
Product Image	 <p>The diagram shows a mobile radiological diagnostic device with a long arm. The arm is labeled 'Arm' and the problem location is indicated by a red dot and labeled 'Problem location'.</p>
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie

Director General







نتقدم بثقة
Moving Forward
with Confidence



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 23 dated 29/01/2023 Regarding NCMDR FSCA of Monnal T60 ventilators from (mfr: Air Liquide Medical Systems).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information




PADC
المديرية العامة للصيدلة والرقابة الدوائية
Directorate General of Pharmaceutical
Affairs & Drug Control



ص.ب: ٣٩٣ مسقط - الرمز البريدي: 100 - هاتف: ٢٢٣٥٧١١١ - فاكس: ٢٢٣٥٨٤٨٩

P.O. Box: 393 Muscat - Postal Code: 100 - Tel: 22357111 - Fax: 22358489

Twitter: @dgpa_dc Email: dg-padc@moh.gov.om





Kingdom of Saudi Arabia
Saudi Food & Drug Authority

Medical Devices Sector

قطاع الأجهزة الطبية

- Home
- Published FSNs/Recalls
- About NCMDR
- Contact Us
- FAQ
- Login

NCMDR

National Center for Medical Devices Reporting


المركز الوطني لبلاغات الأجهزة والمنتجات الطبية

NCMDR Recall

Reference Number: mdprc 008 01 23 000

[Back](#)

Date submitted: 1/15/2023

Manufacturer:	Air Liquide Medical Systems.
Device Type:	Monnal T60 ventilators
Description:	Ventilators
Medical Device Identifier:	Monnal T60 range: References: KA010000 - KA013700 - KA017114 - KA017115 Monnal T60 Advanced range: References: KA017119 - KA017124 - KA017122 - KA017127 - KA017128 - KA017129 - KA017130.
Reason of Field Safety Corrective Action:	Air Liquide Medical Systems has assessed three situations concerning the settings of ventilation parameters for the Monnal T60 range that could pose a risk, despite the existing software safety features: Situation 1: Possibility of having an Fio2 setpoint applied that is different to the setpoint displayed after using the 100% O2 function. Situation 2: Inheritance of setpoints. Situation 3: Oxygen therapy.
Remedy Action:	Air Liquide Medical Systems requests installation of the latest available software version correcting the listed situations during the next maintenance of the device (preventive or corrective), and at the latest within one year.
Athorized Representative/Importer/Distributor:	Bio Standards
Report Source:	NCMDR
Source Ref. Number:	A7E50F3BE32B9
SFDA Comments:	SFDA urges all healthcare providers that have devices subjected to this safety alert to contact the company.
Attachments:	 FSN R2218602 Air Liquide Medical Systems.pdf

[View History](#)

Copyright © 2008 Saudi Food and Drug Authority. All rights reserved.

