



نقدم بثقة
Moving Forward
With Confidence

رؤية عُمان
2040
Oman Vision

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 73 dated 19/4/2023 Regarding NCMDR Recall of Syringe from (mfr: Saudi Mais for Medical Products).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information



PADC
المديرية العامة للصيدلة والرقابة الدوائية
Directorate General of Pharmaceutical
Affairs & Drug Control



ص.ب: ٣٩٣ مسقط - الرمز البريدي: ١٠٠ - هاتف: ٢٢٣٥٧١١١ - فاكس: ٢٢٣٥٨٤٨٩

P.O. Box: 393 Muscat - Postal Code: 100 - Tel: 22357111 - Fax: 22358489

🌐 dgpa_dc Email: dg-padc@moh.gov.om



Circular No. 73 / 2023

نتقدم بيقين
Moving Forward
with Confidence

رؤية عمان
2040
Oman Vision

28 -09-1444 H

19 -04-2023

Recall of Syringe from Saudi Mais for Medical Products.

Source	NCMDR - National Center Medical Device Reporting- SFDA. https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=15&rid=19503
Product	Syringe.
Description	Single-Use Devices.
Manufacturer	Saudi Mais for Medical Products.
Local agent	National Pharmacy LLC.
The affected products	LOT: 221122.
Reason	Reason of recall is leakage because cap is not secured and cracked.
Action	1. Stop marketing and recall all affected products from the market. 2. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om



Dr. Mohammed Hamdan Al Rubaie
Director General