



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 178 dated 28/9/2022 Regarding NCMDR Field Safety Corrective Action of Artis system from (mfr: SIEMENS).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 178 / 2022

بنقدم بثقة
Moving Forward
with Confidence



2 -03-1444 H

28 09-2022

Field Safety Corrective Action of Artis system from SIEMENS.

Source	NCMDR- National Center for Medical Devices Reporting- SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=10&rid=17266
Product	Artis system.
Description	Interventional fluoroscopic x-ray system.
Manufacturer	SIEMENS.
Local agent	Muscat Pharmacy & Stores LLC.
The affected products	System IDs: Artis Q / Q.zen / pheno / icono Model numbers: 10848280, 10848281, 10848282, 10848283, 10848353, 10848354, 10848355, 10848460, 10849000, 11327600, 11327700, 11328100
Reason	Siemens Healthcare is informing users of a potential issue with the Artis system dedicated error detection mechanism. In rare cases of failure of the error detection mechanism, it is not possible to release X-rays until the system resets. If this issue occurs, the system message "No X-ray, tube too hot!" is shown without an additional audible sound and the operator is unable to release X-rays. A system shutdown is required to enable the release of X-ray.
Action	1. Siemens are advising customers that a hardware modification will be performed to correct the issue and they will be contacted in order to schedule an appointment. In the interim, customers are also advised that a shutdown and restart of the system should recover the normal operation of the system. 2. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie
Director General

