



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No...3.3..... dated 17/2/22 Regarding NCMDR recall of RAMY 3ml Luer Lock Syringe with Needle from (mrf: Alshifa Medical Products Co. (CJS)).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 33 / 2022


16 -07-1443 H

17 -02-2022

بمقدم بثقة
Moving Forward
with Confidence

رؤية عمان
2040
Vision

Recall of RAMY 3ml Luer Lock Syringe with Needle from Alshifa Medical Products Co. (CJS).

Source	NCMDR-National Center for Medical Device Reporting https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=4&rid=16018
Product	RAMY 3ml Luer Lock Syringe with Needle.
Description	Hypodermic Syringes.
Manufacturer	Alshifa Medical Products Co. (CJS).
Affected	RAMY 3ml Luer Lock Syringe with needle G23x 1 ¼" Lot# 17793 Ref: F30302312221
Local Agent	Salalah Medical Supplies Manufacturing Co. LLC.
Reason	Some 3ml Luer Lock Syringe with Needle G23x1 ¼" reported to be blocked when use.
Action	1. Immediately discontinue use and quarantine affected syringes (Lot# 17793). 2. Contact the local agent for remedial action.
Product Image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control contact E-mail: Med-device@moh.gov.om



AS
/ Dr. Mohammed Hamdan Al Rubaie

DIRECTOR GENERAL