



To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES
Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)
Director General of Engineering Affairs, MOH
Director General of Royal Hospital
Director General of Khoula Hospital
Director General of Medical Supplies (MOH)
Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)
Hospital Director (Al Nahda Hospital)
Hospital Director (Al Massara Hospital)
The Head of Medical Services in SQU Hospital
The Head of Medical Services in Royal Oman Police
The Head of Medical Services in Ministry of Defence
The Head of Medical Services in The Diwan
The Head of Medical Services in The Sultan's Special Force
The Head of Medical Services in Internal Security Services
The Head of Medical Services in Petroleum Development of Oman
The Head of Medical Services in LNG Oman
ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 206 dated 14/11/2022 Regarding NCMDR recall of 3MTM Steri-Drape™ Surgical Drapes from (mfr: 3M Health Care).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DGPA&DC
- Director of Pharmacovigilance & Drug Information Dept, DGPA&DC
- Director of Drug Control Department, DGPA&DC
- Director of Pharmaceutical Licensing Department, DGPA&DC
- Director of Central Quality Control Lab., DGPA&DC
- Supdt. of Central Drug Information





Circular No. 2061/2022


بمقدم بثقة
Moving Forward
with Confidence



20 -04-1444 H

14 -11-2022

Recall of 3MTM Steri-Drape™ Surgical Drapes from 3M Health Care.

Source	NCMDR- National Center for Medical Devices Reporting- SFDA https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=4&rid=17319
Product	3MTM Steri-Drape™ Surgical Drapes.
Description	Drape, surgical, general-purpose, single use.
Manufacturer	3M Health Care.
Local agent	Waleed Pharmacy & Stores L.L.C
The affected products	Catalogue numbers 1000; Lot number: 33HECP Catalogue numbers 1010; Lot number: 33HJNT, 33KM9K, 33KPJY Catalogue numbers 1020; Lot number: 33JXEM, 33KL5W, 33KLMN Catalogue numbers 1021; Lot number: 33H9HW Catalogue numbers 1060; Lot number: 33K9TF Catalogue numbers 1061; Lot number: 33KJTC.
Reason	The liner on the adhesive component of the affected drapes is difficult to remove without damaging the product and may render the product unusable. In addition, there is an increase in reported adhesive related skin injuries for these affected lots.
Action	1. Please identify the affected product listed above and immediately cease use of affected models and lots of 3M™ Steri-Drape™ Surgical Drapes. 2. Please discard all remaining affected product listed above per facility procedures. 3. Contact the local agent for remedial action.
Product image	
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: Med-device@moh.gov.om

Dr. Mohammed Hamdan Al Rubaie
Director General

