Sultanate of Oman Ministry of Health Drug Safety Center Muscat



سلطنة عُمـان وزارة الصحـة مركز سلامة الـدواء مسقط

To:

THE DIRECTOR GENERAL OF HEALTH SERVICES IN ALL GOVERNORATES

Commanding Officer, Armed Forces Hospital (Al Khoudh & Salalah)

Director General of Engineering Affairs, MOH

Director General of Royal Hospital

Director General of Khoula Hospital

Director General of Medical Supplies (MOH)

Director General of Pvt. Health Est. Affairs (to kindly arrange distribution to all Pvt. Hospitals)

Hospital Director (Al Nahda Hospital)

Hospital Director (Al Massara Hospital)

The Head of Medical Services in SQU Hospital

The Head of Medical Services in Royal Oman Police

The Head of Medical Services in Ministry of Defence

The Head of Medical Services in The Diwan

The Head of Medical Services in The Sultan's Special Force

The Head of Medical Services in Internal Security Services

The Head of Medical Services in Petroleum Development of Oman

The Head of Medical Services in LNG Oman

ALL PRIVATE PHARMACIES & DRUG STORES

After Compliments,

Please find attached our Circular No 87 dated 28/4/2025 Regarding SFDA Recall of Posterior Tibialis Tendon - Anterior Tibialis Tendon from (mfr: MTF Musculoskeletal Transplant Foundation).

Copy to:

- Director, Office of H.E. The Undersecretary for Health Affairs
- Director of Medical Device Control, DSC
- Director of Pharmacovigilance & Drug Information Dept, DSC
- Director of Drug Control Department, DSC
- Director of Pharmaceutical Licensing Department, DSC
- Director of Central Quality Control Lab., DSC
- Supdt. of Central Drug Information





Sultanate of Oman Ministry of Health Drug Safety Center Muscat



سلطنة عُمــان وزارة الصحــة مركـز سلامـة الــدواء

29 -10-1446 H 28 -04-2025

Circular No. 87/2025

Recall of Posterior Tibialis Tendon - Anterior Tibialis Tendon from MTF Musculoskeletal Transplant Foundation.

Source	SFDA- Saudi Food & Drug Authority. https://ade.sfda.gov.sa/Fsca/PublishDetails/351
Product	Posterior Tibialis Tendon - Anterior Tibialis Tendon.
Manufacturer	MTF Musculoskeletal Transplant Foundation.
Local agent	Imtnan Company for Medical Supplies.
The affected products	Posterior Tibialis Tendon: Product Code: 400340, 400340 Serial Number: 00423090891045, 00423090891050 Anterior Tibialis Tendon: Product Code: 400335 Serial Number: 00423090891049
Reason	This voluntary market withdrawal is being initiated due to a notification received by the CDC related to sterility concerns.
Action	 Please return the unused affected products to the local distributor. Contact the local agent for remedial action.
comments	Healthcare professionals are encouraged to report any adverse events Suspected to be associated with the above device or any other medical device to Department of Medical Device Control through the E-mail: vigilance-md@moh.gov.om

Ph. Ibrahim Nasser Al Rashdi **Director General**



